

## Declaration of Conformity

**Company** Smartbox Assistive Technology Ltd.  
**SRN** GB-MF-00000008343  
**Address** Ysobel House  
Malvern  
Worcestershire  
WR14 1JJ  
United Kingdom  
**Phone** +44 (0) 1684 578868  
**Web** www.thinksmartbox.com

**Authorised Representative** REHAVISTA GmbH  
Konsul-Smidt-Str. 8c  
28217 Bremen  
Germany  
**Phone** 0421 - 989 628 - 0

We, Smartbox Assistive Technology Ltd. under our sole responsibility, declare that the product listed below:

**Type of Product** Communication Aid  
**Description** A dedicated communication aid supplied with Grid 3 software and service package. The device can be used with and has been tested with switch, pointer, and eye gaze access methods. It can be supplied with or without these accessories.  
**Product Name** Grid Pad 16  
**Model Number** GP16A  
**UDI (Unique Device Identifier)** 5060446901465

The object of this declaration is a Class I Medical Device and is in conformity with the following EU harmonised legislation:

**2002 No 618** UK Medical Device Regulations  
**2017/745** The EU Regulation on Medical Devices (MDR)  
**2011/65/EU** ROHS  
**2012/19/EU** WEEE  
**1907/2006** REACH

The following harmonized and/or unharmonized standards and technical specifications have been applied:

**ISO 14971:2019** Application of risk management to medical devices  
**EN 60601-1:2006/A1:2013** Medical electrical equipment - Part 1: General requirements for basic safety and essential performance  
**IEC / EN 60601-1-2:2015** Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests  
**EN 50581:2012** Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

This declaration is sign on behalf of Smartbox Assistive Technology Ltd by:



Simon Poole  
27 / 02 / 2025  
Group Chief Technology Officer